

WADSWORTH CENTER
NEW YORK STATE DEPARTMENT OF HEALTH
ENVIRONMENTAL LABORATORY APPROVAL PROGRAM

Proficiency Test Report

Lab Id: 11642 USGS NATIONAL WATER QUALITY LABORATORY
 EPA Lab Code: POB 25046-MS 407-BLDG 95
 CO00009 DENVER, CO-80225-0046
 UNITED STATES
 (303) 236-2000
 Director: MR. MERLE SHOCKEY

Shipment Date : 18-Jul-2005
 Closing Date : 01-Sep-2005
 Score Date : 21-Sep-2005

This report may contain data that are not covered by the NVLAP accreditation.

** Indicates NVLAP accredited analyte evaluated using the USEPA's National Standards for Water Proficiency Testing Studies Criteria Document.
 NVLAP Lab Code 200387-0.

Shipment: 285 Non Potable Water Chemistry

Analyte Name	Units	Sample ID	Method	Result	Mean/ Target *	Warning Limits	Acceptance Limits	Score
Sample: Non Potable Water Base/Neutrals								
1,2,4-Trichlorobenzene EPA Code: N/A	ug/L	8536	USGS O-3116-87	29.9	39.0		10.6 - 67.3	Satisfactory 74 passed out of 75 reported results.
2,4-Dinitrotoluene EPA Code: N/A	ug/L	8536	USGS O-3116-87	43.4	48.5		21.0 - 76.1	Satisfactory 69 passed out of 71 reported results.
2,6-Dinitrotoluene EPA Code: N/A	ug/L	8536	USGS O-3116-87	< 2.3	< 20.0 *			Satisfactory 68 passed out of 71 reported results.
2-Chloronaphthalene EPA Code: N/A	ug/L	8536	USGS O-3116-87	< 1.0	< 20.0 *			Satisfactory 71 passed out of 71 reported results.
4-Chlorophenylphenyl ether EPA Code: N/A	ug/L	8536	USGS O-3116-87	< 1.2	< 25.0 *			Satisfactory 71 passed out of 71 reported results.
4-Bromophenylphenyl ether EPA Code: N/A	ug/L	8536	USGS O-3116-87	26.4	28.6		12.0 - 45.2	Satisfactory 69 passed out of 70 reported results.
Acenaphthene EPA Code: N/A	ug/L	8536	USGS O-3116-87	29.9	32.5		16.7 - 48.3	Satisfactory 74 passed out of 77 reported results.
Anthracene EPA Code: N/A	ug/L	8536	USGS O-3116-87	113	121		61.2 - 181	Satisfactory 78 passed out of 79 reported results.
Acenaphthylene EPA Code: N/A	ug/L	8536	USGS O-3116-87	62.5	72.0		35.4 - 109	Satisfactory 76 passed out of 78 reported results.
Bis(2-chloroethyl)ether EPA Code: N/A	ug/L	8536	USGS O-3116-87	138	118		42.3 - 193	Satisfactory 64 passed out of 72 reported results.
Bis(2-chloroethoxy)methane EPA Code: N/A	ug/L	8536	USGS O-3116-87	< 1.0	< 10.0 *			Satisfactory 71 passed out of 71 reported results.
Bis(2-chloroisopropyl)ether EPA Code: N/A	ug/L	8536	USGS O-3116-87	< 1.0	< 30.0 *			Satisfactory 70 passed out of 71 reported results.